


FORMTECTOR XP-308B

Your Authorized New Cosmos Distributor:

Minerva Industrial & Trading Pte Ltd

T: 6848 5155 E: sales@minerva-intra.com W: www.minerva-intra.com.sg



Features

- No need to replace reagents for each measurement.
- Select either mg/m³ or ppm as the displayed measurement mode for the measurement time.
- Select either the 30-min or 10-min measurement mode for the measurement time.
- Fully automatic operation simply by pushing operation buttons at the site of measurement.
- Random access to measurement data after measurement (up to 32 values).
- Reduced influence of VOCs through the use of DNPH filter.

Applications

- Indoor air control.
- Housing material control.
- Checking remodeling byproducts.

Specifications

Model	XP-308B
Substance Detected	Formaldehyde
Detection Principle	Electrochemical cell (Influence of VOCs reduced by DNPH filter.)
Detection Range	0.01 to 0.38mg/m ³ (Service range: 0.39 to 3.00mg/m ³) 0.01 to 0.30ppm (Service range: 0.31 to 3.00ppm)
Sampling Method	Extractive
Measurement Time	30min/10min measurement mode (selectable)
Power Source	Six C dry cells or AC adaptor (9VDC)
Battery Life *1	30-min measurement mode : Approx. 50 measurement operations (by manganese dry cells)(10-min measurement mode : Approx. 100 measurement operations)
Filter Life *2	30-min measurement mode : 50 measurement operations (10-min measurement mode : 100 measurement operations)
Operating Temperature	5 to 40°C
Display	Digital numbers, Message, Bar graph
Dimensions	W175 × H140 × D86mm
Weight	Approx. 2.5kg
Standard Accessories	3 DNPH Filters, 6 C manganese dry cells, Soft case, Sampling tube, Ten filter elements, Tweezers
Optional Accessories	AC adaptor, External output cable, Data reception software

*1 Battery life may vary with storage conditions, conditions of use, storage period, battery manufacturers, etc.

*2 These values are for an average formaldehyde concentration per measurement of 0.01mg/m³. The battery life will be shortened at higher concentrations.

Considerations for Use

- Measured value can be influenced if aldehyde group other than formaldehyde, like acetaldehyde, coexists.
- Cigarette smoke or alcohol like liquor will cause measurement errors.
- Sudden change in environment, such as temperature and humidity, will cause measurement errors.